

# BIOCOMPATIBILITY CERTIFICATE

2<sup>nd</sup> edition (translation)

**Testmaterial:** Bacoban<sup>®</sup> 2

**Supplier:** sarastro GmbH; Zum Schacht 7  
D-66287 Quierschied-Göttelborn

**Studies performed:** The following studies were performed in order to determine the biocompatibility of the product according to ISO 10993-1:

## CYTOTOXICITY

## SENSITISATION (LLNA<sup>1)</sup>)

## DERMAL IRRITATION


**Results:** The product did not show any effect in the Sensitisation and Dermal Irritation study. A cytotoxic effect was obtained. Due to the intended use as a disinfectant, the effect could be anticipated.

- 1) Local Lymph Node Assay
- 2) tested as sarastro<sup>®</sup> surf.pro-life

**BSL BIOSERVICE Scientific Laboratories GmbH Munich**

Behringstraße 6

D-82152 Planegg

  
Dr. Ingrid Haist  
Biological Safety Testing

Date: July 22, 2005

